

നമ്പർ - എ1/1557/2022

ജില്ല ആശുപത്രി, തിരുർ
തീയതി - 05/12/2022

ഫോൺ - 04942972000

ദർഘാസ് പരസ്യം

മലപ്പുറം ജില്ല പഞ്ചായത്തിന്റെ 2022-2023 വാർഷിക പദ്ധതി നമ്പർ - എസ്0495/2023 ൽ ഉൾപ്പെടുത്തി മോഡുലാർ ഓപ്പറേഷൻ തിയേറ്ററിലേക്ക് ആവശ്യമായ Multiparameter monitor ... അഞ്ച് എണ്ണം വിതരണം ചെയ്യുന്നതിന് അംഗീകരണ വിതരണകാരിൽ നിന്നും ദർഘാസ് ക്ഷണിച്ച് കൊള്ളുന്നു. വിശദാംശം ചുവടെ ചേർക്കുന്നു.

Sl no		Quantity
1	Multiparameter monitor with following accessories, 1. Reusable adult and paediatric SpO2 probe with cable 2. ECG leads (5 lead) with cable 3. Adult and paediatric NIBP cuff (all sizes) with cable. 4. Temperature probe.	05
2	Monitor stand	05

Extra items to be quoted,

SL NO	ITEMS	QUANTITY
1	Reusable adult SpO2 probe with cable	04
2	Reusable paediatric SpO2 probe with cable	04

അടങ്കൽ തുക - 496600/-

ദർഘാസ് ഫോമിന്റെ തുക - 1000/- + ജി.എസ്.ടി



നിരത ദ്രവ്യം - 5000/-

ദേശസാൽകൃത ബാങ്കിൽ നിന്നും എടുത്ത ഡി.ഡി ആയിരിക്കണം

ദർഘാസ് ഫോറം വിതരണം ചെയ്യുന്നത് - 12/12/2022 , 12.00 പി.എം

ദർഘാസ് ഫോറം വിതരണം ചെയ്യുന്ന അവസാന തീയതി - 27/12/2022 , 12.00 പി.എം

ദർഘാസ് തുറക്കുന്നത് - 27/12/2022 , 3.00 പി എം

സ്വീകർത്താവ് - ഡയറക്ടർ പി & ആർ ഡി തിരുവനന്തപുരം

പകർപ്പ് -

1. ആരോഗ്യ വകുപ്പ് ഡയറക്ടറുടെ വെബ്സൈറ്റ്
2. സെക്രട്ടറി ജില്ല പഞ്ചായത്ത് മലപ്പുറം
(നോട്ടീസ് ബോർഡിൽ പതിപ്പിക്കുന്നതിന്)

സ്പെസിഫിക്കേഷൻ ഇതോടൊപ്പം ഉള്ളടക്കം ചെയ്യുന്നു.

ദർഘാസ് കവറിന്റെ പുറത്ത് MULTIPARAMETER MONITOR... ഓപ്പറേഷൻ തിയേറ്റർ ജില്ല ആശുപത്രി , തിരുർ എന്ന് കാണിക്കണം

ദർഘാസ് സമ്പന്നമായ എല്ലാ സർക്കാർ വ്യവസ്ഥകളും ബാധകമായിരിക്കും. ദർഘാസ് സമ്പന്നമായ കൂടുതൽ വിവരങ്ങൾക്ക് പ്രവൃത്തി സമയങ്ങളിൽ ഓഫീസു മായി ബന്ധപ്പെടാവുന്നതാണ്.



സൂപ്രണ്ട്

Superintendent
DISTRICT HOSPITAL
TIRUR - 676 104

Specifications

1. Should be able to monitor 6 channel ECG from 5 lead ECG, SPO2, NIBP, Respiration Rate and temperature.
 2. Should be portable with carrying handle.
 3. Should have touch screen TFT display with at least 12 inches or higher with at least 6 waveforms and numeric display simultaneously.
 4. Should have Lithium ion battery with 3 hours battery back up.
 5. Should have keys for quick access to main functions.
 6. Should have adult, pediatric and neonatal modes.
 7. Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia
 8. Recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.
 9. NIBP can be taken on manual/auto/stat modes.
 10. Each Multiparameter Monitor Should provide following accessories:
 - i. Reusable adult 5 lead ECG cable set – 1 no.
 - ii. Reusable adult and pediatric SPO2 finger probes – 1 each.
 - iii. Adult and pediatric NIBP cuff of different size – 1 each
 - iv. Reusable Skin Temperature probe - 1
- Rate of Reusable adult and pediatric SPO2 finger probes should be quoted Separately
11. Should work on 200-240V AC/50Hz with inbuilt rechargeable battery.
 12. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate/ STQC S certificate or valid detailed electrical and functional safety test report from ERTL.
Copy of the certificate / test report shall be produced along with the technical bid to verify at the time of Technical Bid opening.
 13. Should have display perfusion index
 14. Monitor should have networking facility with bidirectional & bed to bed communication.
 15. Monitor stand should be supplied along with the machine as per the following specification
 - i. Monitor stand -extruded Aluminum, powder coated.

