

SPECIFICATION

Tender No.1/2013-14

Blood Lancets

Disposable and Sterile

Batch No., Date of manufacture and date of expiry should be specified in the invoice at the time of supply.

75% shelf life at the time of supply.

Tender No.2/2013-14

Micro slides

Ground edge, lint free and fog free

Size: 75/76 mm X wide: 25/26 mm

Thickness: 1.2 – 1.35 mm

Tender No.3/2013-14

Temephos : 50%EC IS: 8498/1977

1. Notary attested copy of the CIB Registration Certificate
2. Notary attested copy of the Manufacturing License.
3. Notary attested copy of the Distribution License issued by the Department of Agriculture, Kerala
4. Batch Number, Date of manufacture and Date of expiry should be specified in the invoice at the time of supply.
5. 75% shelf life at the time of supply
6. Test report from the accredited laboratory should be submitted at the time of supply.



Addl. Director of Health Services (Medical)

Tender No.4/2013-14

PYRETHRUM EXTRACT 2%

1. Notary attested copy of the CIB Registration Certificate
2. Notary attested copy of the Manufacturing License.
3. Notary attested copy of the Distribution License issued by the Department of Agriculture, Kerala
4. Batch Number, Date of manufacture and Date of expiry should be specified in the invoice at the time of supply.
5. 75% shelf life at the time of supply
6. Test report from the accredited laboratory should be submitted at the time of supply.

Tender No.5/2013-14

Mechanical Aspirator (Battery operated mechanized aspirator)

Mechanical aspirator with Extension Tube 5-6

Acrylic, Rubber stopper and Collection bottle assembly : 15 Nos.

Additional requirement:

- | | | |
|-------------------------------|---|---------|
| 1. Extension tube | : | 75 Nos. |
| 2. Rubber stopper | : | 75 Nos. |
| 3. Collection bottle assembly | : | 75 Nos. |



Addl. Director of Health Services (Medical)

Tender No.6/2013-14

Spinal decompression system

- Should provide facility for de compression for pelvic as well as cervical.
- Pull pattern should replicate manual therapy.
- Should have pre programmed user friendly protocols.
- Independent timer, progression, traction and regression.
- Should have facility to vary speed of traction.
- Should provide de compression in both prone as well as supine positions.
- System should have table and appropriate accessories.
- Should have facility for pelvic tilt in a table.
- Should have the facility to change the pull angle.
- Should provide gradient pulls, oscillations, and progressive range of motion.
- Necessary belting to be supplied with the system.
- Should store patient data, library includes anatomical images, common pathologies and belting techniques.
- Mains power 100-240v.50-60hz.

Warranty :

The entire equipment and its accessories should have **warranty of three years** from the date of installation /commissioning. Ternderer should also certify that, goods supplied under this contract are new, unused and have no defects arising from design, materials, workmanship etc. If the down time due to idling of equipment for want of repair/maintenance under warranty period exceeds 5% (55 days in the case of 3 year warranty) of the total warranty, period of warranty shall be extended accordingly.

CMC : The rate and terms of CMC **for a period of 5 years** after the warranty period should be clearly specified.

Safety Certificate: Should have safety certificate from a competent authority CE/FDA(US)/STQC CB Certificate/STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be submitted along with the Tender.



Addl. Director of Health Services (Medical)

Tender No.7/2013-14

Vibrator

- Unit mounted on castor stand with accessories tray.
- Output speed 15-60 cycles/ second.
- Power output 90 VAC through 240 VAC 50/60hz.
- Electronic speed control.
- Able to adapt to various body surfaces.
- Should have variable speed output controls.

Motor 1/17 horse power, 24VDC.

Warranty :

The entire equipment and its accessories should have **warranty of three years** from the date of installation /commissioning. Tenderer should also certify that, goods supplied under this contract are new, unused and have no defects arising from design, materials, workmanship etc. If the down time due to idling of equipment for want of repair/maintenance under warranty period exceeds 5% (55 days in the case of 3 year warranty) of the total warranty, period of warranty shall be extended accordingly.

CMC : The rate and terms of CMC for a period of 5 years after the warranty period should be clearly specified.

Safety Certificate: Should have safety certificate from a competent authority CE/FDA(US)/STQC CB Certificate/STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be submitted along with the Tender.



Addl. Director of Health Services (Medical)

Tender No.8/2013-14

Shoulder CPM (Continuous Passive Motion)

- Hand held pendent control with chip card.
- Facility to use left or right arm.
- Anatomically correct alignment.
- Unit should switch to the opposite direction when patient resistance increases target set value.
- Should allow full shoulder range.
- Adjustable ergonomic chair for adequate shoulder stabilization.

Warranty :

The entire equipment and its accessories should have **warranty of three years** from the date of installation /commissioning. Ternderer should also certify that, goods supplied under this contract are new, unused and have no defects arising from design, materials, workmanship etc. If the down time due to idling of equipment for want of repair/maintenance under warranty period exceeds 5% (55 days in the case of 3 year warranty) of the total warranty, period of warranty shall be extended accordingly.

CMC : The rate and terms of CMC for a period of 5 years after the warranty period should be clearly specified.

Safety Certificate: Should have safety certificate from a competent authority CE/FDA(US)/STQC CB Certificate/STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be submitted along with the Tender.

Tender No.9/2013-14: This tender is cancelled.



Addl. Director of Health Services (Medical)